

Calibration Certificate

Model No:
Serial No:
Temperature: 22.50 °C
Relative Humidity: 35 %

Calibration Date: 26 January 2006
Certificate No:
Procedure: QSIW-2420 REV. A

Customer Information:

Initial Condition (As Received):

- In Tolerance
- Out of Tolerance
- Repairs Needed
- Not Applicable (New Product)

Final Condition (As Shipped):

- In Tolerance

- Keithley Instruments, Inc. certifies that the above instrument meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology (NIST), or other national standards, and meets the technical requirements of ANSI/NCSL Z540-1-1994.
- The quality system used by Keithley Instruments, Inc. is registered to ISO 9001:2000.
- This calibration is a direct comparison of the unit under test to the listed reference standards and did not involve any sampling plans to complete. No allowance has been made for the instability of the test device due to use, time, etc. Such allowances would be made by the customer as needed.
- This calibration certificate shall not be reproduced except in full without the written approval of Keithley Instruments, Inc.

Calibration Standards Used

<u>Manufacture</u>	<u>Model Number</u>	<u>Control Number</u>	<u>Cal Date</u>	<u>Due Date</u>
HEWLETT PACKARD	3458A	6364	12/20/2005	3/20/2006
Fluke	5700A	7444	1/9/2006	4/9/2006

Technician: Tom Horvath

Approved By: Helga A. Alexander

Helga A Alexander
Metrology Team Leader